

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N: 130113

Supervisor/Operator Performing the Verification Procedure:

Name: Kenneth G. Cox ID #: 3527 Date: April 2, 2009

A

Agency: Skagway Police Department Phone #: 907 983-2232

Instrument Location Skagway Police Department, 1<sup>st</sup> & State Street Skagway, Alaska 99840

Alco S/N X173010 Target Value .082 High Pressure 400

Alco Test Value Average .080 .081  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature

(OVER)

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 15 day of May 2009.

Carolyn M. Noland (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



APR 06 2009

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130113

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113

APRIL 02, 2009 ✓

OPERATOR'S NAME:

COX, KENNETH G

OPERATOR'S NUMBER: 3527

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: SKG1

CASE/REPORT: 2009-0534

TEST TYPE: U

ALCO TARGET VALUE: .002 ✓

ALCO S/N: X173010

### — BREATH ANALYSIS —

.002 ADJUSTED FOR 30.09 in

ALCO TARGET .002 06:13

BLANK TEST .000 06:15

INTERNAL STANDARD VERIFIED 06:15

ALCO TV 30.09 in .000 ✓ 06:15

BLANK TEST .000 06:16

SUBJECT SAMPLE .000 06:17

BLANK TEST .000 ✓ 06:17

ALCO TV 30.09 in .001 06:18

BLANK TEST .000 06:19

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130113

APRIL 02, 2009 ✓

TIME 06:03

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 40c

BREATH TUBE: 40c

BAROMETER: 30.07 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEF  
GHIJKLMNOPQRSTUVWXYZ[\]^\_`abcdefghijklmnopqrstuvwxyz{|}~